

# Real-time PCR detection of *Streptococcus pneumoniae* in respiratory specimen BAL, whole blood, cerebrospinal fluid and urine using the Luminex ARIES® platform



Rocio B. Damiano, Subathra Marimuthu, Megan E. Greenwell and Leslie A. Wolf  
Division of Infectious Diseases, University of Louisville, Louisville, Kentucky, USA

## INTRODUCTION

- Community-acquired pneumonia (CAP) and complications such as bacteremia and meningitis due to *Streptococcus pneumoniae* infection still occurs in at risk populations, despite the availability of effective vaccines.
- Laboratory confirmation of *S.pneumoniae* remains challenging in cases of CAP despite advances in blood culture techniques and the availability of nucleic acid amplification tests.
- In our experience, current microbiological methods and diagnostic assays leave up to 80% of in-hospital patients without an identified pathogen causing pneumonia, potentially leading to misuse of antibiotic.
- In this study, we are targeting a portion of the *S. pneumoniae* specific gene (autolysin, *lytA*) for detection. Using RT-PCR MultiCode® Technology, we tested four different clinical samples types: bronchoalveolar lavage (BAL), whole blood (WB), cerebrospinal fluid (CSF) and urine to cover the various presentations of invasive pneumococcal infections.




Figure 1. Testing scheme

## INSTRUMENT

- The ARIES® is an in vitro diagnostic medical device for detection of nucleic acids by fluorescence based PCR.
- The *lytA* and MHV 2 primer pool were added to the ReadyMix® tubes, attached to the cassettes. Next, the clinical samples were added to the cassette and loaded in the ARIES®.
- The ARIES® generates PCR amplification and melt curves for the target, with resulting Ct values calculated for both the sample processing control and the target.



## DISCUSSION AND CONCLUSIONS

- The significance of this method validation is to provide a rapid and sensitive test for assessing *S. pneumoniae* infection using four common specimen types.
- A CLIA-validated diagnostic test for invasive pneumococcal infections benefits the local healthcare community by providing quality test results within a few hours of specimen receipt.
- In addition, the ARIES® platform minimizes hands on time for laboratory technologists and allows for validation of laboratory developed tests

## MATERIALS AND METHODS

- S. pneumoniae* primer pairs were designed to include one fluorescent reporter labeled primer with an MultiCode® isoC on the 5' end with an unlabeled primer, both obtained from IDT. The primer sequences are /56-FAM//iMe-isodC/A CGC AAT CTA GCA GAT GAA GCA and CTC CCT GTA TCA AGC GTT TTC GGC
- The sample processing control primer murine hepatitis virus 2 (MHV 2), ReadyMix® and cassettes were purchased from Luminex. The cassette with ReadyMix® contains all components needed to extraction of nucleic acid and perform PCR.
- Sensitivity (Limit of Detection, or LOD), precision, accuracy and stability for BAL, WB, CSF and urine were tested to validate the assay to meet CLIA requirements. Waste clinical samples were used and spiked with *S. pneumoniae* when necessary (Figure 1).

## RESULTS

**Analytical Sensitivity:** *S.pneumoniae* organisms were serially diluted ten-fold from 1 x 10<sup>4</sup> to 1 x 10<sup>-1</sup> CFU/mL for these experiments. Results are from three replicate, independent experiments. The different sample types LOD for *S.pneumoniae* are shown in the following table.

Sample type	LOD (CFU/ml)	<i>lytA</i> Ct value Avg(SD)
BAL	1 X 10 <sup>2</sup>	35.23 (0.50)
WB	1 X 10 <sup>2</sup>	36.00 (0.70)
CSF	1 X 10 <sup>2</sup>	33.97 (0.47)
Urine	1 X 10 <sup>1</sup>	37.13(1.63)

**Precision:** Reproducibility of the resultant Ct value did not change over the course of the different testing period by different operators. The standard deviation (SD) for *lytA* Ct values were less than 2.5, on all specimen types.

**Accuracy:** Ten previous *lytA* positive were like wise positive with the assay and Ten negative samples for all different specimen types were also negative using *lyt A* assay, giving a 100 % accuracy for all sample types

**Analytical stability:** Spiked positive specimens were stored in refrigerator for 8 days. All specimens appeared stable at 4°C temperature for up to 8 days.

## FUNDING

This research received no specific grant from any funding agency in the public, commercial, or not-for-profit sectors.

## CONTACT INFORMATION

Leslie A. Wolf, PhD  
Infectious Diseases Laboratory  
Room 104 MDR Building  
511 South Floyd Street  
University of Louisville  
Louisville, KY 40292 (40202 for courier)

Voice 502-852-1523  
FAX 502-852-1512  
Email [leslie.wolf@louisville.edu](mailto:leslie.wolf@louisville.edu)  
Web [www.uoflidlab.com](http://www.uoflidlab.com)